

INCORPORATED

ANALYSIS RESEARCH CONSULTING

2233 Jesse Street
Los Angeles, CA 90023
Tel : (323) 266-3960
Fax: (323) 266-39862310 N. W. 55th Court #123 & 124
Fort Lauderdale, Florida 33309-2743
Tel : (954) 486-2233
Fax: (954) 486-2886

March 21, 2007

World Centric
2211 Williams Street
Palo Alto, CA 94306**File Number:207515**

Attn: Clare Cordero

Phone: (650) 223-4087

Fax: (866) 850-9732

Re: One sample of Corn Starch Utensils (forks, knives and spoons) for Pesticide Residue analysis
Received:02/23/07**RESULTS****(A) Organochloride scan****ECD SCAN:** (Rtx-5/Rtx-CLP@ 150°C for 3 min. then 10°C /min to 290°C)

No organochloride pesticides detected.

Organochloride pesticides detection limit= 0.005 to 1.0 ppm

(B) Organonitrogen scan**TSD SCAN:**(Rtx-5/Rtx-50@ 150°C for 3 min then 10°C/min to 280°C)

No organonitrogen pesticides detected.

Organonitrogen pesticides detection limit = 0.03 to 3.0 ppm

(C) Organophosphate scan**FPD"P" SCAN:**(Rtx-50@140°C for 2 min then 10°C/min to 290°C)



No organophosphate pesticides detected.

Organophosphate pesticides detection limit = 0.01 to 1.0 ppm

(D) HPLC Scan

<u>Compound</u>	<u>Concentration(ppm)</u>	<u>Detection Limit(ppm)</u>
N-Methyl Carbamates	None Detected	0.01

No N-Methyl Carbamates detected.

Reported by
ANRESKO, Inc.
Hing-Man Mang
Senior Chemist
Thomas Ahenguah
Senior Chemist

G:\2007\February\207515 World Centric Corn Starch Utensils Full MRA