



ANALYSIS • RESEARCH • CONSULTING

April 15, 2009

World Centric  
2121 Staunton Court  
Palo Alto, CA 94306

**File Number: 320090613**

Attn: Aseem Das

Phone: (650) 283-3797

Phone: (866) 850-9732

Re: One composited sample (Wheat Straw Plate, Wheat Straw Clamshell and Wheat Straw Bowl) for full multi-residue analysis

Received:03/31/09

Method of extraction: Pesticide Analytical Manual (PAM) Vol I Section 302-E4

### RESULTS

#### (A) Organochloride Scan

ECD SCAN: (Rtx-5/Rtx-35@ 150°C for 3 min. then 10°C /min to 290°C )

No organochloride pesticides detected.

Organochloride pesticides detection limit= 0.005 to 0.30 ppm

(B) Organonitrogen Scan (Rtx-5/Rtx-50@ 150°C for 1 min then 10°C/min to 280°C )

No organonitrogen pesticides detected.

Organonitrogen pesticides detection limit = 0.03 to 1.0 ppm

#### (C) Organophosphate Scan

FPD"P" SCAN:(Rtx-5@ 100°C for 2 min then 10°C/min to 290°C )

No organophosphate pesticides detected.


Organophosphate pesticides detection limit = 0.01 to 0.30 ppm

#### (D) HPLC Scan

<u>Compound</u>	<u>Concentration(ppm)</u>	<u>Detection Limit(ppm)</u>
N-Methyl Carbamates	None Detected	0.01

No N-Methyl Carbamates detected.

Reported by  
ANRESKO, Inc.

  
Hing-man Mang  
Senior Chemist

  
Slate Carrigg  
Chemist

G:/2009/March/320090613 World Centric Wheat Straw Plate, Wheat Straw Clamshell and Wheat Straw Bowl full mra.wpd